



## Certificate of Analysis

Baird-Parker Agar  
Catalog Number: B02-102  
Lot Number: 201020022310  
Expiration: October 2023

3651 Clipper Mill Rd.  
Baltimore, MD 21211  
Phone (410) 467-9983  
Fax (410) 467-5088

### Physical And Chemical Tests

<u>Test Parameter</u>	<u>Specification</u>	<u>Result</u>
DCM Appearance	Homogeneous, free flowing, beige	Homogeneous, free flowing, beige
PPM Appearance*	Opaque, canary yellow	Opaque, canary yellow
pH at 25 °C (6.0% solution)	7.0 +/- 0.2	7.01

### Microbiological Tests

<u>Microorganisms</u>	<u>Growth*</u>	<u>Reactions*</u>
<i>Bacillus subtilis</i> ATCC 6633	Inhibition	N/A
<i>Enterococcus faecalis</i> ATCC 29212	Excellent Black col. w/o halo	Lipase & Lecithinase -
<i>Escherichia coli</i> ATCC 25922	Inhibited	N/A
<i>Proteus mirabilis</i> ATCC 12453	Fair Brown col. w/o halo	Lipase & Lecithinase -
<i>Staphylococcus aureus</i> ATCC 6538	Excellent Black col. w/ halo	Lipase & Lecithinase +
<i>Staphylococcus aureus</i> ATCC 25923	Excellent Black col. w/ halo	Lipase & Lecithinase +
<i>Staphylococcus epidermidis</i> ATCC 12228	Excellent Grey-Bk col. w/o halo	Lipase & Lecithinase -

\* The following results were obtained from type cultures after incubation at a temperature of 37°C and observed after 24-48 hours. Media prepared with Egg Yolk Tellurite Enrichment.

### Laboratory Results: Satisfactory

Country of Manufacture:  
Testing performed by:  
Completed on:  
QC Release by:

USA  
A.Kazmi/ H.McGinnis  
11-12-20  
Abida Kazmi