



# Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079  
 TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

12/10/2020 21:53:32 GMT  
 Page 1 of 1

Customer Name: **Bel-Art Products-NJ** Processing Facility: **Salem** Work Order # **2785729**  
 P.O.# **MD61424** Sales Order # **2646216**  
**23.5-50.0 kGy** BelArt, Gamma Treatment Irradiation Date/Time: **12/10/2020 14:19:00 GMT**

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
103.000	100	CA	369440010(30)	STRLWR,SMPL SPOON,1 TSP,PS,10/	J347439	NA
104.000	20	CA	369441010	STERILEWARE®,SMPLSPOON,1TSP	J347495	NA
105.000	58	CA	369461010	STRLWR,SMPLSPOON,3TSP,DBLBAG	J347497	NA
106.000	3	CA	378730001	TRANSPETTE,PIPETTOR,LDPE,8CHAN	J346515	NA
107.000	90	CA	369460000	STERILEWARE,SAMPLING SPOON,3 T	J346767	NA
108.000	300	CA	369460010	Tablespoon 1.0	J347239	NA
109.000	20	CA	367400004	STRLWR,ECON SMPE SPN,PS 8ML	J346382	NA
110.000	20	CA	369300050	STERILEWARE®,SAMPLING SPATULA,	J345850	NA
119.000	40	CA	369022002(20)	STERILWARE,SCOOP,RED,PS,2 OZ	J346962	NA
120.000	300	CA	369020006(20)	STERILEWARE®,METAL DETECT	J346762	NA
				SCOO		
	<b>951</b>	<b>CA</b>	<b>Total</b>			

### Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	23.5 kGy	50.0 kGy	26.8 KGY	Pass	SBARBIERI	12/10/2020 17:33:29 GMT
		Reason Code Test				STEVEN BARBIERI	
450.00	Maximum Dose	23.5 kGy	50.0 kGy	43.9 KGY	Pass	SBARBIERI	12/10/2020 17:33:38 GMT
		Reason Code Test				STEVEN BARBIERI	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: **HEATHER ZMIJEWSKI**  
 Reason: **Work Order Completions**

Date: **12/10/2020 21:50:30 GMT**