



Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079
 TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

05/24/2021 12:06:18 GMT
 Page 1 of 1

Customer Name: Conbar
 P.O.# N/A
 10.0 - 40.0 kGy

Processing Facility: Salem
 Conbar, Gamma Treatment

Work Order # 2885703
 Sales Order # 2769561

Irradiation Date/Time: 05/22/2021 08:03:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	1	EA	Scoops 343-1099	343-1099-BL-40-48-75	2152505	5665S coliwasa
102.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8882-24 scoop/5665S
103.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8882-24/2400-25S
104.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8864-24 scoop/5327S
105.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8832-25 scoop
106.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8832-25 scoop
107.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8832-25 scoop
108.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8832-25 scoop
109.000	1	EA	Scoops 343-1099	148-637-BL-47-35-55	2152505	8832-25 scoop
	9	EA	Total	343-1099-BL-40-48-75	2152505	2440-25S/2405-25S

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	10.0 kGy	40.0 kGy	12.8 KGY	Pass	JNEWMAN Jeffrey Newman	05/22/2021 10:25:09 GMT
450.00	Maximum Dose	Reason Code Test 10.0 kGy	40.0 kGy	27.5 KGY	Pass	JNEWMAN Jeffrey Newman	05/22/2021 10:25:20 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: AMY L. EDWARDS
 Reason: Work Order Completions

Date: 05/24/2021 12:06:03 GMT