



Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079
TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

05/24/2021 12:06:18 GMT
Page 1 of 1

Customer Name:	Conbar	Processing Facility:	Salem	Work Order #	2885703
P.O.#	N/A			Sales Order #	2769561
	10.0 - 40.0 kGy	Conbar, Gamma Treatment	Irradiation Date/Time:		05/22/2021 08:03:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	1	EA	Scoops 343-1099	343-1099-BL-40-48-75	2152505	5665S coliwasa
102.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8882-24 scoop/5665S
103.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8882-24/2400-25S
104.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8864-24 scoop/5327S
105.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8832-25 scoop
106.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8832-25 scoop
107.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8832-25 scoop
108.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152505	8832-25 scoop
109.000	1	EA	Scoops 343-1099	343-1099-BL-40-48-75	2152505	8832-25 scoop
	9	EA	Total			2440-25S/2405-25S

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	10.0 kGy	40.0 kGy	12.8 KGY	Pass	JNEWMAN	05/22/2021 10:25:09 GMT
450.00	Maximum Dose	Reason Code Test 10.0 kGy	40.0 kGy	27.5 KGY	Pass	Jeffrey Newman	05/22/2021 10:25:20 GMT
		Reason Code Test				JNEWMAN	
						Jeffrey Newman	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: **AMY L. EDWARDS**
Reason: **Work Order Completions**

Date: 05/24/2021 12:06:03 GMT